

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIIMS-BLS/ECRlab/2024/NIQ/201

Dated: 01/10/2024

NOTICE INVITING QUOTATION

Supply of Electronic Pipettes in the Department of Endocrinology and Clinical Immunology and Rheumatology Lab

Quotation as a PASSWORD protected PDF file is invited from registered stockists/ distributors/ manufacturers having GST No. for the Supply of Electronic Pipettes in the Department of Endocrinology and Clinical Immunology and Rheumatology Lab as per the detailed specifications (Annexure-A). Submit quotation as a PASSWORD_protected PDF **“Quotation No. AIIMS-BLS/ECRlab/2024/NIQ/01** for the Supply of Electronic Pipettes in the Department of Clinical Immunology and Rheumatology Lab containing both the “Technical Bid” and “Price Bid” (in two separate files). The files may be submitted so as to reach on or before 22/10/2024 up to 11:00 A.M in ECR lab, 1st floor C-Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. Bidders will have to quote for all the items mentioned in the quotation. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee.

For the purpose of technical evaluation, the bidder is required to submit following documents.

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Turnover for last two years (2021-22 & 2022-23).
4. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
5. Full description, make and quantity of the item must be mentioned in the quotation.
6. Technical Evaluation will be based on total value, for the participants having the compliance to the required technical specifications.
7. A technical compliance sheet need to be provided for every item included in the bid. The participants can be invited for a physical demonstration by the user department.

In case the Technical Committee rejects a sample (if asked for) or the bid is “Not Qualified” on technical grounds, the financial bid in respect of that item will not be considered.


Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail. Should have comprehensive warranty for at least 1 years or more.

Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other terms and conditions will be as follows: -


1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. **Date of opening** of NIQ will be communicated to the participant well in advance.
12. Taxes at other government levies will be paid extra as applicable.
13. Venue for sample verification (if asked for) and quotation opening shall be in ECR lab, 1st floor C-Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact Dr Ravi Kumar Sharma, Faculty In charge ECR lab, 8427472362, E-mail rheumatology.iiimsbilaspur@gmail.com



Faculty In Charge
ECR lab
AIIMS-Bilaspur
H.P.

Requirement for electronic pipettes:

S. No	Capacity	Channels	Quantity
1.	10-100 μl	1	1
2.	10-100 μ l	12	1
3.	100-1000 μ l	1	2
4.	5-50 μ l	1	1
5.	0.5-10 μ l	1	1
6.	0.5-10 μ l	12	1

Dr. Devender Bainwa, D.M. (Clinical Immuno. & Rheumatology) AIIMS
M.D. (MEDICINE) AIIMS, New Delhi
Assistant Professor, Dept. of Internal Medicine
AIIMS, Bilaspur, H.P. (DMCIR/09133)


Associate Professor
Endocrinology & Metabolism
AIIMS Bilaspur (H.P.)


Dr Ravi Kumar Sharma
Assistant Professor, Clinical
Immunology & Rheumatology
AIIMS Bilaspur (H.P.)


Specifications for electronic pipettes


1. Electronic pipette Single channel electronically adjustable 10-100 μ L

Specification	Requirement
Pipette type	1 channel
Micropipette with spring loaded tip cone	Yes
Autoclavable	Yes
Locking Mechanism	Yes
Digital Volume Indicator	Yes
Capacity to dispense multiple times after one aspiration	Yes
Volumetric range	10-100 μ L or similar
ISO/CE/BIS certification with identifiers	Required
Warranty	3 years warranty with one free annual calibration
Material of Piston	Chemical resistant
Ease and reproducibility	Must deliver reliable and repeatable results, regardless of the operator
Memory operations	Should be able to store at least 5 programs
Service and calibration reminders	Compliance required

2. Multichannel electronically adjustable 10-100 μ L 1 Quantity

Specification	Requirement
Electronic Multi Channels Pipette 8 Channel And 12 Channels	8 channels
Micropipette with spring loaded tip cone	Yes
Autoclavable	Yes
Locking Mechanism	Yes
Digital Volume Indicator	Yes
Capacity to dispense multiple times after one aspiration	Yes
Volumetric range	10-100 μ L or similar
ISO/CE/BIS certification with identifiers	Required
Warranty	3 years warranty with one free annual calibration
Material of Piston	Chemical resistant
Microwell Plate Tracker	Pipette should support this feature
Ease and reproducibility	Must deliver reliable and repeatable results, regardless of the operator
Memory operations	Should be able to store at least 5 programs
Service and calibration reminders	Compliance required


 Associate Professor
 Endocrinology & Metabolism
 AIIMS Bilaspur (H.P.)


 Dr Ravi Kumar Sharma
 Assistant Professor, Clinical
 Immunology & Rheumatology
 AIIMS Bilaspur (H.P.)


Dr. Devender B...
 M.D. ...
 Assistant ...
 AIIMS ...

Annexure 8: Specifications for electronic pipettes digitally adjustable 100-1000uL, Quantity-2

Specification	Requirement
Pipette type	1 channel
Micropipette with spring loaded tip cone	Yes
Autoclavable	Yes
Locking Mechanism	Yes
Digital Volume Indicator	Yes
Capacity to dispense multiple times after one aspiration	Yes
Volumetric range	100-1000 μ L
ISO/CE/BIS certification with identifiers	Required
Warranty	3 years warranty with one free annual calibration
Material of Piston	Chemical resistant
Ease and reproducibility	Must deliver reliable and repeatable results, regardless of the operator
Memory operations	Should be able to store at least 5 programs
Service and calibration reminders	Compliance required


Dr. Devender Bainwa, D.M. (Clinical Immunology & Rheumatology) IP
M.D. (MEDICINE) AIIMS Bilaspur, H.P.
Assistant Professor (Dept. of Internal Medicine)
AIIMS, Bilaspur, H.P. (DMCR/2013)



Associate Professor
Endocrinology & Metabolism
AIIMS Bilaspur (H.P.)


Dr Ravi Kumar Sharma
Assistant Professor, Clinical
Immunology & Rheumatology
AIIMS Bilaspur (H.P)

Annexure 9: Specifications for electronic pipettes digitally adjustable 10-50uL

Specification	Requirement
Pipette type	1 channel
Micropipette with spring loaded tip cone	Yes
Autoclavable	Yes
Locking Mechanism	Yes
Digital Volume Indicator	Yes
Capacity to dispense multiple times after one aspiration	Yes
Volumetric range	10-50 µL or similar
ISO/CE/BIS certification with identifiers	Required
Warranty	3 years warranty with one free annual calibration
Material of Piston	Chemical resistant
Ease and reproducibility	Must deliver reliable and repeatable results, regardless of the operator
Memory operations	Should be able to store at least 5 programs
Service and calibration reminders	Compliance required



Associate Professor
Endocrinology & Metabolism
AIIMS Bilaspur (H.P.)



Dr Ravi Kumar Sharma
Assistant Professor, Clinical
Immunology & Rheumatology
AIIMS Bilaspur (H.P)

Dr. Devender Bairwa, D.M. (Clinical Immuno. &
 Rheumatology)
 M.D. (MEDICINE) (Post Graduate)
 Assistant Professor (Clinical Immunology & Rheumatology)
 AIIMS, Bilaspur, H.P. (DNC-209133)

Annexure 10: Specifications for electronic pipettes digitally adjustable 0.5-10uL

Specification	Requirement
Electronic Multi Channels Pipette 8 Channel And 12 Channels	12 channels
Micropipette with spring loaded tip cone	Yes
Autoclavable	Yes
Locking Mechanism	Yes
Digital Volume Indicator	Yes
Capacity to dispense multiple times after one aspiration	Yes
Volumetric range	0.2-10 µL or similar
ISO/CE/BIS certification with identifiers	Required
Warranty	3 years warranty with one free annual calibration
Material of Piston	Chemical resistant
Microwell Plate Tracker	Pipette should support this feature
Ease and reproducibility	Must deliver reliable and repeatable results, regardless of the operator
Memory operations	Should be able to store at least 5 programs
Service and calibration reminders	Compliance required


Associate Professor
Endocrinology & Metabolism
AIIMS Bilaspur (H.P.)


Dr Ravi Kumar Sharma
Assistant Professor, Clinical
Immunology & Rheumatology
AIIMS Bilaspur (H.P)

Dr. Devender Bainwa, D.M. (Clinical Immuno &
Rheumatology) & F.
M.D. (MEDICINE) Bilaspur, Jharkhand
Assistant Professor (Department of Medicine)
AIIMS, Bilaspur, H.P. (CHARTERED)

Specifications for electronic pipettes digitally adjustable 0.5-10 μ L

Specification	Requirement
Pipette type	1 channel
Micropipette with spring loaded tip cone	Yes
Autoclavable	Yes
Locking Mechanism	Yes
Digital Volume Indicator	Yes
Capacity to dispense multiple times after one aspiration	Yes
Volumetric range	0.2-10 μ L or similar
ISO/CE/BIS certification with identifiers	Required
Warranty	3 years warranty with one free annual calibration
Material of Piston	Chemical resistant
Ease and reproducibility	Must deliver reliable and repeatable results, regardless of the operator
Memory operations	Should be able to store at least 5 programs
Service and calibration reminders	Compliance required



Associate Professor
Endocrinology & Metabolism
AIIMS Bilaspur (H.P.)

Dr. Devender Bairwa, D.M. (Clinical Immuno. &
Rheumatology) JIPMER
M.D. (MEDICINE) AIIMS New Delhi
Assistant Professor (Dept. of Internal Medicine,
AIIMS, Bilaspur, H.P. (DMC/R/05133))



Dr. Ravi Kumar Sharma
Assistant Professor, Clinical
Immunology & Rheumatology
AIIMS Bilaspur (H.P.)