

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIIMS-BLS/DENTISTRY/2024/NIQ/05

Dated: 14/12/2024

NOTICE INVITING QUOTATION

Supply of Surgical micromotor with console for the Department of Dentistry

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for supply Surgical micromotor with console for the Department of Dentistry as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as "Quotation No. AIIMS-BLS/DENTISTRY/2024/NIQ/05 for the for supply of Equipment and Consumables for the Department of Dentistry due date of opening 29 / 12/2024" containing both the "Technical Bid" and "Price Bid" may be submitted so as to reach on or before 28/12/2024 up to 11:00 A.M in the R.No B 749, Department of Dentistry 4th floor B block AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh. PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee and bidders who may wish to be present on the same day at 11:30 A.M.


For the purpose of technical evaluation, the bidder is required to submit following documents:

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during 2021-22 and 2022-23 for the similar products.
4. Previous year Income Tax Return (AY- 2020-21, 2021-22 and 2022-23).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical Evaluation will be based on total value wise.
8. Bidder must submit sample for each item along with bid as per specification for technical evaluation.

In case the Technical Committee rejects a sample or the bid is "Not Qualified" on technical grounds, the financial bid in respect of all the items will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.



Medical Superintendent
चिकित्सा अधीक्षक
AIIMS, Bilaspur (H.P.)
अ.मा.आ.स. बि.नातपुर (हि.प्र.)

Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other terms and conditions will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in Department of Dentistry 4th floor B- Block, AIIMS Bilaspur, Kothipura, Himachal Pradesh PIN- 174001.
13. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact Dr. Dinesh Kumar Verma, Faculty In charge, Department of Dentistry, E-mail office.ms@aiimsbilaspur.edu.in


 Faculty In Charge
 Department Of Dentistry
 AIIMS-Bilaspur
 H.P.
 AIIMS, Bilaspur (H.P.)
 ज.भा.आ.स. बिलासपुर (हि.प्र.)

Annexure A

Department of Dentistry

Sl. No	Name	Technical specifications	Qty
1.	SURGICAL MICROMOTOR WITH CONSOLE	<ol style="list-style-type: none"> 1. Should be US FDA or European CE with 4 digit notified body no. or BIS certified 2. Should be a Motor-Brushless induction motor 3. The Control Unit should have easy control of speed, motor direction, torque and irrigation pumps. 4. It should work within 50-60 Hz or 100-240V 5. The irrigation pumpage should not be more than 120ml/minute. 6. The Max rpm should be atleast 40000 rpm 7. Motors should have high consistent torque 8. The micromotor and handpiece should be with a illumination (LED Based/Fiber optic) 9. The MicroMotor with Cable should be autoclavable upto 134°C. 10. The MicroMotor should come with an intra coupling ("E" coupler) so that allows for universal attachment between motors and handpieces regardless of make 11. The Micromotor with Intra coupling should be upgradeable such that Microcompass saw and Micro sagittal saw can be attached in future 12. The Control unit should also come with a foot pedal 13. The Footpedal should have protection proof class of atleast IP x 5. 14. The Footpedal should have on and off for irrigation pump and have change over function for clock wise and anticlockwise rotation. 15. Should include Cleaning & Lubricating spray 16. All the accessories for basic functions should be provided. 17. One Straight surgical handpiece (1 :1), autoclavable, compatible with 2.35mm dental shank bur. 18. The bidder will give rate list of all possible spares, accessories & consumables if any, as part of financial bid. If price of any spare is not mentioned & is required for repair in life time of equipment/instrument, then the firm will be obliged to give it free throughout life cycle of the equipment. 19. Unit price for 200 irrigation tubing should be quoted separately. Price should be valid for warranty and CMC period. 20. 15 Nos. of burr should be provided along with the unit. 	1

Medical Superintendent
 P.D. Singh
 AIIMS, Bhopal (H.P.)
 31.01.2018