

Corrigendum for Supply of Portable Cardiac Biomarkers (Troponin I Or Troponin T) Machine for the Department of CTVS at AIIMS Bilaspur, H.P.


File No. AIIMS-BLS/CTVS/NIQ/2026-01


AMENDMENTS IN THE TECHNICAL SPECIFICATIONS

S.No. A), 3)	Reaction time for Troponin T or Troponin I : less than 15 minutes for Quantitative results with or without 5 minutes for Qualitative positive results
S.No. A), 5)	Test measuring time should be less than 20 minutes.
S.No. A), 6)	Test strips should have room temperature stability for immediate testing. If cold storage required then test strip should be ready to use immediately after cold storage. No warming to room temperature. (No waiting period).
S.No. A), 9)	Keypad and/or touch screen facility
S.No. A), 13)	Sample size should not be more than 250 micro liters and premeasured pipette to be provided with kit.
S.No. A), 7c)	Dosing modes-triggered or continuous or manual
S.No. A), 15)	There should be possibility of transferring data via QR code, Wi-Fi, USB or data management software
S.No. A), 17)	The test strip should have inbuilt internal quality control and/or manufacture should also have external liquid control to ensure the integrity of test-strip
S.No. F), 1)	Warranty: 2 years

LAST DATE EXTENDED TO 12/05/2026 till 2pm.


Dr. Vikas Kumar
Associate Professor
Dept. of Cardiothoracic Surgery
AIIMS Bilaspur (H.P.)
Regd. No. 10222209


Dr. Navneet Sharma
Associate Professor
Dept. of Burns & Plastic Surgery
AIIMS Bilaspur (H.P.)
Regd. No. 639/20


Medical Superintendent
विक्रमाजीजी
AIIMS Bilaspur (H.P.)
Bilaspur (H.P.)