

23

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIIMS-BLS/(G)/2025/NIQ/

Dated: 07/07/2025

NOTICE INVITING QUOTATION

Supply of Oxygen Cylinder (Type-A), & N₂O Cylinder (Type-A) for the Department of Anaesthesiology and Critical Care, AIIMS ,Bilaspur (H.P.)

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for supply of Oxygen Cylinder (Type-A) & N₂O Cylinder (Type-A) for the Department of Anaesthesiology and Critical Care, AIIMS Bilaspur (H.P.) as per the detailed specifications (Annexure-A). Sealed Quotations containing both Technical and financial bid separately (separate envelop) duly super subscribed at the top of the envelope as **“Quotation No. AIIMS-BLS/ANAESTH/25-69 with end date** for the for supply of Oxygen Cylinder (Type-A) & N₂O Cylinder (Type-A) for the Department of Anaesthesiology & Critical Care. Containing both the “Technical Bid” and “Price Bid” may be submitted so as to reach on or **before 24/07/2025** up to 11:00 A.M in the Procurement section, E- Block, Ground floor, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee.

For the purpose of technical evaluation, the bidder is required to submit following documents:

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during 2022-23 and 2023-24 for the similar products.
4. Previous year Income Tax Return (AY- 2021-22, 2022-23 and 2023-24).
5. Self-declaration that the firm is not debarred by any Govt. Organisation or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical/Financial Evaluation will be based on total value wise.

In case the Technical Committee rejects a sample or the bid is “Not Qualified” on technical grounds, the financial bid in respect of all the items will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.

22
Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other **terms and conditions** will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Bidder is required to quote rate of 100% items.
6. Part supply will not be accepted.
7. No revision in rate (on higher side) will be accepted at any stage.
8. The firm shall not assign or sublet the work/job or any part of it to any other firm.
9. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected& accepted. No advance payment will be made at any stage.
10. Taxes at other government levies will be paid extra as applicable.
11. Venue for sample verification (if asked for) and quotation opening shall be in office of Procurement officer, Ground Floor, E-Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001.
12. In case manufacturer participates, dealers will not be allowed or disqualified.
13. For any query, please contact, Faculty In charge Procurement, E-mail hites_store@aiimsbilaspur.edu.in , storeofficer@aiimsbilaspur.edu.in

Faculty In Charge
Procurement Section
AIIMS-Bilaspur
Himachal Pradesh

SPECIFICATIONS

List of Oxygen Cylinders and Connectors required with Mindray Anaesthesia work station

S. No	Item	Quantity
1.	Type A Oxygen cylinder with pin index cylinder valve	40
2.	Type A N ₂ O cylinder with pin index cylinder valve	40

Modified Specifications-

1. Type A Oxygen cylinder with pin index cylinder valve-

- Gas capacity- 0.75 CuM
- Water capacity- 5 liter
- Length- 735 mm approx
- Minimum wall thickness- 4.0 mm, outer diameter- 108 mm
- Should have a neck ring
- Should be supplied with pin Index valve system
- Maximum Working pressure 150 bar
- Purity of Medical Grade Oxygen Gas when tested as per Indian Pharmacopeia (Latest), - Minimum (%) V/V- 99%
- Oxygen gas in the Cylinder should conforms to IP ^{Latest}
- Supply of Medical grade Oxygen gas in CCE approved cylinders only
- ISI marked Valve conforming to IS 3224-2002 (Latest) fitted with the cylinder
- Validity of Cylinder safety certificate- minimum 5 years
- Valve protection caps should be available
- All fittings should be leak proof
- Each cylinder should be engraved with the details viz (a) Serial No of cylinder/Identification, (b) Code of manufacturing, (c) Working pressure, (d) Water capacity in liters, (e) Test pressure in Kg/cm², (f) Year of manufacture and testing, (g) Tare weight of the cylinder
- All the cylinders should be properly painted as per their respective color code, Cylinder No, name of the supplier, name of the gas cylinder and its contents should be clearly indicated on the neck of the cylinder.
- All cylinders must have 5 year PESO filling permission.
- Testing of Cylinder as per standard protocol will be the responsibility of Cylinders supplier.
- Availability of BIS Inspection Certificate.
- History cards of the cylinders as per Gas Cylinder Rules 2016 and rules made thereafter, as amended from time to time should be made available with filled Cylinder supplies, if owner of the Cylinder is supplier
- Availability of Cylinder filling license (Form F) from Chief Controller of Explosives, Department of Petroleum, Government of India
- In case of manufacturer to manufacture and supply Medical grade Oxygen gas in cylinders to IP as per Govt Of India guide lines, availability of valid license (Form 25 & Form 26) from Drug Control or any appropriate authority is mandatory
- Availability of valid Drug License
- On site demonstration to be provided to ensure compatibility with available Mindray anaesthesia work station at AIIMS, Bilaspur.
- Cylinder should be fully filled at the time of supply and the same should be checked and verified.

M. H. P.
 Head of Department
 Anaesthesiology & Critical Care
 AIIMS Bilaspur (H.P.)

[Signature]
 Dr. Sunil Thakur
 Associate Professor
 Deptt. of Anaesthesiology
 AIIMS, Bilaspur (H.P.)
 Regd. No. HPSMC-154

[Signature]
 Assistant Professor
 Department of Anaesthesiology
 and Critical Care
 AIIMS Bilaspur (HP)

- Any leak/defect in cylinders will be resolved/replaced by the supplier only.

2. Type A N₂O cylinder with pin index cylinder valve

- Water capacity- 5 liter
- Length- 735 mm approx.
- Minimum wall thickness- 4.0 mm, outer diameter- 108 mm
- Should have a neck ring
- Should be supplied with pin Index valve system
- Maximum Working pressure 150 bar
- ISI marked Valve conforming to IS 3224-2002 (Latest) fitted with the cylinder
- Validity of Cylinder safety certificate- minimum 5 years
- Valve protection caps should be available
- All fittings should be leak proof
- Each cylinder should be engraved with the details viz (a) Serial No of cylinder/Identification, (b) Code of manufacturing, (c) Working pressure, (d) Water capacity in liters, (e) Test pressure in Kg/cm square, (f) Year of manufacture and testing, (g) Tare weight of the cylinder
- All the cylinders should be properly painted as per their respective color code, Cylinder No, name of the supplier, name of the gas cylinder and its contents should be clearly indicated on the neck of the cylinder
- All cylinders must have 5 year PESO filling permission
- Testing of Cylinder as per standard protocol will be the responsibility of Cylinders supplier
- Availability of BIS Inspection Certificate.
- History cards of the cylinders as per Gas Cylinder Rules 2016 and rules made thereafter, as amended from time to time should be made available with filled Cylinder supplies, if owner of the Cylinder is supplier
- Availability of Cylinder filling license (Form F) from Chief Controller of Explosives, Department of Petroleum, Government of India
- In case of manufacturer to manufacture and supply Medical grade Oxygen gas in cylinders to IP as per Govt Of India guide lines, availability of valid license (Form 25 & Form 26) from Drug Control or any appropriate authority is mandatory
- Availability of valid Drug License
- On site demonstration to be provided to ensure compatibility with available Mindray anaesthesia work station at AIIMS, Bilaspur.
- Cylinder should be fully filled at the time of supply and the same should be checked and verified.
- Any leak/defect in cylinders will be resolved/replaced by the supplier only

[Handwritten Signature]
 Head of Department
 Anaesthesiology & Critical Care
 AIIMS Bilaspur (H.P.)

[Handwritten Signature]
 Dr. Sunil Thakur
 Associate Professor
 Deptt. of Anaesthesiology
 AIIMS, Bilaspur (H.P.)
 Regd. No. HPSMC-154

[Handwritten Signature]
 Assistant Professor
 Department of Anaesthesiology
 and Critical Care
 AIIMS Bilaspur (HP)