

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIIMS-BLS/Pharma/2025/NIQ/01

Dated: 06/01/2025

NOTICE INVITING QUOTATION

Supply of Rotarod Apparatus in the Department of Pharmacology

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for the supply of Rotarod Apparatus in the Department of Pharmacology as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as **“Quotation No. AIIMS-BLS/Pharma/2025/NIQ/01”** for the for supply of Rotarod Apparatus to the Department of Pharmacology **due date of opening 21/01/2025** containing both the “Technical Bid” and “Price Bid” may be submitted so as to reach on or before 21/01/2025 up to 11:00 A.M in the Department of Pharmacology, Academic Block, 2nd floor, AIIMS Bilaspur, Kothipura, Himachal Pradesh PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee and bidders who may wish to be present on the same day at 11:30 A.M.

For the purpose of technical evaluation, the bidder is required to submit following documents:

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during 2022-23 and 2023-24 for the similar products.
4. Previous year Income Tax Return (AY- 2021-22, 2022-23 and 2023-24).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Physical demonstration of equipment will be part technical evaluation.
8. Technical Evaluation will be based on total value wise.
9. Bidder must submit sample for each item along with bid as per specification for technical evaluation.

In case the Technical Committee rejects a sample or the bid is “Not Qualified” on technical grounds, the financial bid in respect of all the items will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a

minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.

Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other **terms and conditions** will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in the Department of Pharmacology, Academic Block, 2nd floor, AIIMS Bilaspur, Kothipura, Himachal Pradesh PIN- 174001.
13. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact at e-mail ID office.pharmac@aiimsbilaspur.edu.in

Dr. Deepti Chopra
Professor & Head
Department of Pharmacology

HOD/Faculty In Charge
Department of Pharmacology
AIIMS, Bilaspur. (H.P)

Rotarod Apparatus (Rat & Mice- acceleration)

Rotarod test is widely used to evaluate drug effects on motor coordination, balance and motor learning in small laboratory animals (mice and rat).

Specifications

1. Should be usable for both mice and rat
2. Should be at least five compartment model.
3. Should have the lane width which can be easily adjusted for mice and rat
4. Should have detection of fall by plate type sensor.
5. Should have individual lane timers in the range of 0-990 seconds with resolution of at least 0.1s
6. Should have an electronic rotor speed adjustment with a constant speed in the range of 0.1-75 rpm at least.
7. Should have provision for easy replacement of rotor.
8. Should have at least 3 different modes- Normal mode, 2 different acceleration modes.
9. Should also have reverse- forward direction selection mode.
10. Should have the lane width of at least 50mm for mice and & 70mm for rat.
11. Should have the falling height at least 220mm for mice & 230 mm for rat.
12. Should have rotor diameter of at least 30 mm.
13. Parameters recorded: Distance travelled in centimeters, falling RPM and time spent on rod.
14. Should be provided with password-protected software for data collection.
15. Should have touch screen display with functions such as password protected system, Cut-off time, Auto Experiment Number, Password protected delete mode.
16. Software should have provision for entering experiment title, comments, graphical presentation of data and assignment of test ID to compartment for animal groups.
17. Software should have PC/laptop connectivity
18. Software with transfer of data to excel, report generation in excel or pdf, & calibration report generation facility.
19. Should have software with provision to add title, comment, graphical presentation and data can be exported to PC/laptop.
20. Should provide 5-year warranty with 5-year CMC.
21. Should be ISO certified.

(H)

Shreenakshi

U. S. S. AA

ad