



अखिलभारतीयआयुर्विज्ञानसंस्थान, बिलासपुर
हिमाचलप्रदेश-१७४००१
ALL INDIA INSTITUTE OF MEDICAL SCIENCES, BILASPUR
HimachalPradesh-174001

<http://aiimsbilaspur.edu.in>



Ref: AIIMS-BLS/Pharma/2025/NIQ/07

Dated- 13/01/2025

CORRIGENDUM

Supply of Analgesiometer Tail Flick (Digital) in the Department of Pharmacology

With reference to NIQ No. AIIMS-BLS/Pharma/2025/NIQ/07 published on 06.01.2025, all interested parties are notified that the tender for Supply of Analgesiometer Tail Flick (Digital) in the Department of Pharmacology, AIIMS Bilaspur (H.P.) has been amended as follows:

1. Analgesiometer Tail Flick (Digital) specifications as per Annexure.

All other terms and conditions of the tender remain same as per order no. AIIMS-BLS/Pharma/2025/NIQ/07 dated 06.01.2025.

Accepted

Head of Department
Department of Pharmacology,
AIIMS Bilaspur (H.P.)


Analgesiometer Tail Flick (Digital)


Analgesiometer Tail Flick (Digital) is used to perform rapid & precise screening of analgesic property of a drug on small laboratory animals.

Specifications

1. Should have lamp illumination system, preferably shutter operated.
2. Should provide heating with halogen lamp.
3. Should have Digital Lamp Intensity controller with Lamp Intensity in the range between 1-99% in around 50 steps.
4. Should have both automatic and manual mode for detecting tail flick.
5. Should have remote switch to control the test start / stop to allow rapid Hands-free experiments
6. Should have a built-in timer activated by an external switch to allow the precise measurement of the reaction time (resolution up to 0.1 Sec).
7. Should have provision to add maximum sequence numbers, auto experiment number and animal sex.
8. Should give precise and reproducible results.
9. Should have cut off timer to avoid damage to the animal.
10. Should provide PC Connectivity with software.
11. Should provide software with calibration and should report generation feature
12. Should provide software with provision to add title, comment, graphical presentation and data can be converted in excel or pdf and exported to a PC/laptop.
13. Should provide three mice & three rat restrainers.
14. Product must be CE & ISO certified.
15. Should provide 5-year warranty with 5-year CMC.




13/01/2025


13/01/25