

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIIMS-BLS/Endo/2023/NIQ/ 1

Dated: 22/9/2023

NOTICE INVITING QUOTATION

Supply of Various Equipment/Instruments for Endocrinology Department
AIIMS Bilaspur H.P.

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for the Supply of various equipment/instruments and miscellaneous items for the Department of Endocrinology & Metabolism as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as "**Quotation No. AIIMS-BLS/Endo/2023/NIQ/ 1**" for the Supply of various equipment/instruments and miscellaneous items for the Department of Endocrinology & Metabolism **due date of opening _____**" containing both the "Technical Bid" and "Price Bid" (in two separate envelopes) may be submitted so as to reach on or before 17.10.2023 up to 5:00 P.M. in Room No 602/603 Endocrinology OPD, B-Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. Bidders can quote for either all the items or only for the items which are available with him/her. L1 shall be calculated on the basis of price quoted for each individual item/equipment and separate supply orders shall be issued for each item/equipment. Late bids will not be considered. The bids shall be opened in the presence of duly constituted departmental procurement committee .

For the purpose of technical evaluation, the bidder is required to submit following documents.

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during 2022-23 for the similar products.
4. Turnover for last two years (2021-22 & 2022-23).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.

In case the Technical Committee rejects a sample (if asked for) or the bid is "Not Qualified" on technical grounds, the financial bid in respect of that item will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms for each item separately. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.



Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other terms and conditions will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Part supply will not be accepted.
7. No revision in rate (on higher side) will be accepted at any stage.
8. The firm shall not assign or sublet the work/job or any part of it to any other firm.
9. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
10. Taxes at other government levies will be paid extra as applicable.
11. Venue for sample verification (if asked for) and quotation shall be intimated on a later date through email.
12. For any query, please contact Dr. Preyander Singh Thakur, Associate Professor Department of Endocrinology & Metabolism AIIMS Bilaspur H.P., 01978-292575, E-mail office.endo@aiimsbilaspur.edu.in



Dr. Preyander S Thakur
Assoc Prof.
Endocrinology & Metabolism
AIIMS-Bilaspur H.P.

Annexure A

ITEM No.	NAME OF EQUIPMENT/ INSTRUMENT	QTY REQUIRED	SPECIFICATIONS
1.	Semmes Weinstein Monofilaments	50 of each 2g, 4g, 10g	<ul style="list-style-type: none"> • Should be made of Nylon • Sizes : Three : 2g (4.3) , 4g (4.56) & 10 g (5.07) • Number required : 50 each • Monofilaments should rest on bar or handle • Should have universal colour coding
2.	Prader Orchidometer	1	<ul style="list-style-type: none"> • Number Required - 1 • To measure the size of the testicles • Should consists of a chain of 12 testis models having defined volumetric capacities, as follows:- 1, 2, 3, 4, 5, 6, 8, 10, 12, 15, 20, 25ml. • Weight: 0.25kg approx.
3.	Radiographic Atlas of Skeletal Development of the Hand & Wrist	1	<ul style="list-style-type: none"> • Number required - 1 • Publisher – Stanford university Press • Authors : William Walter Greulich & S. Idell Pyle • ISBN -10 : 9780804703987 • Language : English • Edition : 2nd
4.	CGMS	2 with 25 sensors	<ul style="list-style-type: none"> • Should support continuous glucose monitoring - real time • Should be small, comfortable and water resistant • Recorder should be suitable for multi patient use • No needle should stay inside body. • Should work under wide range of temperatures and humidity • Should be internally powered device with continuous mode of operation • Battery life: Should atleast support 7 days of Continuous glucose monitoring (CGM), • Should supply; (i) Recorder/Reader - 2 no, (ii) Sensors – 25 nos (iii) Sensor patch insertion device - as part of sensor kit if use-and-throw type (25 nos) - or separately (04 no if reusable) • Should come with computer connectivity and software. • Should comply with ISO standard for long-term body contact • Should be BIS or European CE or USFDA approved for use in adult and paediatric patients.

			<ul style="list-style-type: none"> • Sensors found to be not working/malfunctioning should be replaced free of cost.
5.	Durometer	1	<ul style="list-style-type: none"> • Number required - 1 • Should be of shore type A • Should have a range of 0-100 degree, 1 division is 1 degree shore-units • Should be of hand held type • Should be analogue/digital type

Du