

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES**  
**Bilaspur, Himachal Pradesh**

No. AIIMS-BLS/DENTISTRY/2024/NIQ/04

Dated: 14/12/2024

**NOTICE INVITING QUOTATION**

Supply of Handpiece cleaning and oiling system for the Department of Dentistry


Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for supply Handpiece cleaning and oiling system for the Department of Dentistry as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as "Quotation No. AIIMS-BLS/DENTISTRY/2024/NIQ/04 for the for supply of Equipment and Consumables for the Department of Dentistry due date of opening 29 /12/2024" containing both the "Technical Bid" and "Price Bid" may be submitted so as to reach on or before 28/12/2024 up to 11:00 A.M in the R.No B 749, Department of Dentistry 4<sup>th</sup> floor B block AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh. PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee and bidders who may wish to be present on the same day at 11:30 A.M. For the purpose of technical evaluation, the bidder is required to submit following documents:

**Technical Bid:**

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during 2021-22 and 2022-23 for the similar products.
4. Previous year Income Tax Return (AY- 2020-21, 2021-22 and 2022-23).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical Evaluation will be based on total value wise.
8. Bidder must submit sample for each item along with bid as per specification for technical evaluation.

In case the Technical Committee rejects a sample or the bid is "Not Qualified" on technical grounds, the financial bid in respect of all the items will not be considered.

**Financial Bid:** The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.


  
Medical Superintendent  
चिकित्सा अधीक्षक  
AIIMS, Bilaspur (H.P.)  
अ.भा.आ.स. बिलासपुर (हि.प्र.)

**Financial Bid submission format for items with quantity required is as under: -**

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

**Other terms and conditions will be as follows: -**

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in Department of Dentistry 4<sup>th</sup> floor B- Block, AIIMS Bilaspur, Kothipura, Himachal Pradesh PIN- 174001.
13. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact Dr. Dinesh Kumar Verma, Faculty In charge, Department of Dentistry, E-mail [office.ms@aiimsbilaspur.edu.in](mailto:office.ms@aiimsbilaspur.edu.in)

  
**Faculty In Charge**  
**Department Of Dentistry**  
**AIIMS-Bilaspur**  
**H.P.**

## Annexure A

## Department of Dentistry

Sl. No	Name	Technical specifications	Qty
1.	Handpiece cleaning and oiling system	<ol style="list-style-type: none"> <li>1. Servicing 4 handpiece in 100 seconds - Requiring approx. 25 seconds per handpiece. Compared with manual lubrication, it should require only 1/3 of the lubricant.</li> <li>2. It should have Built-in chuck lubrication system</li> <li>3. It Purges and lubricates all types of handpieces - Whether contra-angle and straight handpiece, air turbines and air scalars, Endo Contra, Surgical Handpiece. It should offer highly efficient and user-friendly handpiece maintenance &amp; should be fully automatic.</li> <li>4. It should have dual lubrication system - dual lubrication system enables lubrication of two instruments with one spray or, if so desired, the use of two different sprays. One for high speed &amp; one for low speed or surgical hand pieces.</li> <li>5. It should have step air blow function - The air suction system eliminates lubricant mist, so there is less smell by oil.</li> <li>6. It should offer selectable lubrication time option - to select between two lubrication time modes: short oiling mode, long oiling mode and should also have long purging mode for removal of excess oil&amp; other fluids/debris.</li> <li>7. It should have lubricant indicator - To reduce maintenance failure, it should be equipped with a lubrication sensor, signaling when spray is empty with an illuminated light.</li> <li>8. Dimension should be in the range – W 300mm, H 370mm, D 300mm.</li> <li>9. It should have option to use any reputed brand oil spray can.</li> <li>10. It should be supplied with 35 oil cans.</li> <li>11. Manufacturer should offer 05 Years Warranty. SHOULD BE COMPATIBLE WITH JMORITA ALL TYPE OF HANDPIECES</li> </ol>	1

Medical Superintendent  
 अ.भा.आ.स. अधीक्षक  
 AIIMS, Bilaspur (H.P.)  
 अ.भा.आ.स. बिलासपुर (हि.प्र.)