

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIIMS-BLS/Surg.Onco/2025/NIQ/01

Dated: 07/03/25

NOTICE INVITING QUOTATION

**Supply of ATP Monitoring Luminometer under miscellaneous items for the department of
Surgical Oncology**

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for the Supply of ATP Monitoring Luminometer under miscellaneous items to the Department of Surgical Oncology as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as "Quotation No. AIIMS-BLS/Surg.Onco/2025/NIQ/01 for the Supply of ATP Monitoring Luminometer under miscellaneous items Department of Surgical Oncology due date of opening 22/03/25" containing both the "Technical Bid" and "Price Bid" (in two separate envelopes) may be submitted so as to reach on or before 21/04/25 up to 11:00 A.M in the room no.B-716 Department of Surgical Oncology 4th floor B- Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee .

For the purpose of technical evaluation, the bidder is required to submit the following documents.

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during 2023-24 for the similar products.
4. Turnover for last two years (2021-22 , 2022-23 & 2023-24).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical Evaluation will be based on total value wise

In case the Technical Committee rejects a sample (if asked for) or the bid is "Not Qualified" on technical grounds, the financial bid in respect of that item will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.

Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount


Handwritten signature


Handwritten signature


Handwritten signature

Other terms and conditions will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected and accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in room no.B-716 Department of Surgical Oncology 4th floor B- Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. In case manufacturer participates, dealers will not be allowed or disqualified.
13. For any query, please contact Dr. Chitresh Kumar Sharma ,Faculty In charge, department of SurgicalOncology 01978-292575, E-mail :-office.surgon@aiimsbilaspur.edu.in


Faculty In Charge
Surgical Oncology
AIIMS Bilaspur
Dr. CHITRESH KUMAR
Associate Professor
Dept. of Surgical Oncology
AIIMS BILASPUR (H.P.)
Regd. No. UPMC-51678





ANNEXURE- A
General specifications for ATP Luminometer Monitoring System

The desired ATP Monitoring Device should be:

1. Easy to use, light in weight and handheld device.
2. Simple, quick and reliable hand held verification device which should be effective in cleaning verification and should be capable to capture, store and manage data for later analysis and records.
3. The system should provide identification of potential risks and tracking to monitor, allowing for improved hygiene performance.
4. Fast time to result which allows immediate corrective action.
5. Robust to provide consistent results across temperatures (4°C – 45°C).
6. The results obtained by using this device should be in RLU counts.
7. The test is specific for ATP and it is extremely sensitive being able to detect down to 2.5 picograms (pg) of ATP and give results in 15 seconds.
8. Should undergo self calibration mode at every start up.
9. Should have software that stores, manages and quickly retrieves data. (at least 1000 readings)
10. The device should have a digital display.
11. Should have provision for organising user (min 100) and test locations (min 200) as per plans.
12. The system should be able to detect ATP from solid surfaces and liquid samples.
13. Optional Calibration Control Kit to check the efficacy of Instrument.
14. Should have the option of Positive Control Vials option to check efficacy of Swab.
15. Should be made of 100% Re-Cycle able Plastic for Eco-Friendliness.
16. Device should be operational with just two AA rechargeable battery.
17. It should be US FDA/ CE (EU) approved.
18. The vendor should provide 02 years comprehensive warranty and 03 years CAMC after warranty period.
19. Should be supplied with following pre moistened swab tips for reliable and accurate results with min 12 months shelf life and should have liquid stable chemistry and also stable for min 4 weeks at normal temperatures.
 - a) 2500 surface cleaning verification swabs
 - b) 1000 protein residue verification swabs
 - c) 1000 endoscope needle cleaning verification swabs
 - d) 1000 water quality verification swabs.

Devi

B

12