


MIP XS/S SPECIFICATIONS

- A. Nephroscope should have a size of not more than 6 - 8 Fr.
- B. Nephroscope should have a length not less than 22 cm.
- C. Nephroscope should have an automatic pressure control system so that stones once broken, should come out automatically with or Without Pressure irrigation.
- D. Nephroscope should have separate channel for working instruments and irrigation to have optimal irrigation & good visualization.
- E. The degree of view should not be less than 5 -6 degree.
- F. It should have an angled eye piece.
- G. Scope should be supplied with a sheath which should work as an amplatz sheath as well.
- H. Sheath should have an option in length.
- I. 1 sheath along with 1 step dilator should be supplied of the following specifications 11/12 F (15 cm Length).
- J. Dilator should have a central channel for guide wire.


डॉ. संदीप सिंह सेन
Dr. Sandeep Singh Sen
सहायक अध्यापक, शिशु चिकित्सा विभाग
Assistant professor, Department of pediatric surgery
अ.भा.अ.च. बिलासपुर (H.P.)
AIIMS Bilaspur (H.P.)
Regd.No. JK 5172


Dr. Uma Kant Dutt
MS, M.Ch
Assistant Professor
Department of Urology
AIIMS, Bilaspur (H.P.)

1712 St. Denis St.
New York, N.Y.
10019
Tel: (212) 246-1234
Fax: (212) 246-1235