

AIIMS-BLS/NM/2025/NIQ/ 07

Dated: 11/2/2025

NOTICE INVITING QUOTATION

Supply of Shielding Accessories in the Department of Nuclear Medicine

Password protected quotations are invited (Technical and Financial bids in separate files) from the interested parties/ agencies for Shielding Accessories in the Department of Nuclear Medicine as per the detailed specifications (Annexure-A). Both the bids viz. "Technical and Financial" should be sent on the email id: office.nuclearmed@aiimsbilaspur.edu.in before 13-2025 (21 days from publication of NIQ) up to 11:00 A.M in separate PDF files and files to be named as "Technical bid" and "Financial bid" separately. The Financial bid must be password protected. The mail has to be sent with the subject AIIMS-BLS/NM/2025/NIQ/Shielding Accessories. Financial bids will be opened only for the bidders, whose Technical Bids have been found to be acceptable. Late bids will not be considered. The bids shall be opened in the presence of duly constituted Department Purchase Committee.

Technical Bid:

For the purpose of technical evaluation, the bidder is required to submit following documents.

1. Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured since 2020 for the similar products.
4. Turnover for last two years.
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical Evaluation will be based on total value-wise.

In case the Technical Committee rejects a sample (if asked for) or the bid is "Not Qualified" on technical grounds, the financial bid in respect of that item will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technical

Dr. Anshul Sharma MD
Assistant Professor
Department of Nuclear Medicine
All India Institute of Medical
Sciences, Bilaspur, H.P.

Dr. Muninder Kumar
Additional Professor
Radiation Oncology
AIIMS Bilaspur (H.P.)
Reg. No. HPMC 1772/2012

Dr. Priyanka Tallyur
Associate Professor
Radiation Oncology
AIIMS Bilaspur, (HP)-174037
Reg No. HPMC 1772/2012

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

8

qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.

Financial Bid submission format for items with quantity required is as under: -

S.N.	Item Detail	Quantity	Make	Model	HSN code	Unit Price	GST@	Unit Price with Tax	Total Amount

Other terms and conditions will be as follows:

1. Supply will be F.O.R. at AIIMS-Bilaspur.
2. Supply will have to be arranged within 30 days from the issue of supply order. Delayed supplies beyond 30 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples (if submitted) will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in conference hall of Procurement & Stores Department, Basement-1, D-Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001.

Anshul Sharma
Dr. Anshul Sharma, MD
Assistant Professor
Department of Nuclear Medicine
All India Institute of Medical
Sciences, Bilaspur, H.P.

Muninder Kumar
Dr. Muninder Kumar
Additional Professor
Radiation Oncology
AIIMS, Bilaspur (H.P.)
Regd. No. HPMC-798/09

Dr. P. S. Singh
Dr. P. S. Singh
Associate Professor
Radiation Oncology
AIIMS Bilaspur (HP)-174001
Reg No. HPMC 1772/2012


ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

2

13. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact via E-mail: office.nuclearmed@aiimsbilaspur.edu.in


Dr. Anshu Sharma, MD
Assistant Professor
Department of Nuclear Medicine
All India Institute of Medical
Sciences, Bilaspur, H.P.


Dr. Muninder Kumar
Additional Professor
Radiation Oncology
AIIMS, Bilaspur (H.P.)
Reg. No. 1772/2012


Dr. Priyanka Thakur
Associate Professor
Radiation Oncology
AIIMS Bilaspur, (HP)-174037
Reg No. HPMC 1772/2012

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh


ANNEXURE-A


Specifications: Shielding Accessories
Department of Nuclear Medicine

Detailed Specifications:

S. N.	Name of the item	Number Required	Specifications
1.	Lead Vial Shields	5 (Five)	<ul style="list-style-type: none">• Thickness of lead: min 6mm for 4 shields (for SPECT); one shield with 30mm lead thickness (for PET).• Surface: Hygienic and Stainless Steel/Anodized Aluminium• Lid: Should have the ability to be closed in a secure manner (screw/magnetic). Lid while secured should have a separate opening for placement of needle.• Capacity: Should be compatible with elution vials of major suppliers (SDS, Curium, BRIT etc.)
2.	PET Dose Drawing System	1 (One)	<ul style="list-style-type: none">• Shielding: min 30mm Tungsten• Surface: Hygienic and metallic• Lid: Removable with screw and a slot for placement of syringe needle.• Capacity: Should be compatible with supply vials of major 18F-FDG suppliers.• Misc: Should have the tilt functionality with lock in tilted position.
3.	Cap-Tongs for Radioactive Vials	2 (Two)	<ul style="list-style-type: none">• Material: Stainless Steel• Should be compatible with standard vial shields used in PET and SPECT radio pharmacies.• The Tongs should have a lock and finger grip.• The user should be able to take the vials out of the vial shielding without having to lift/tilt the shield.


Dr. Anshul Sharma, MD
Assistant Professor
Department of Nuclear Medicine
All India Institute of Medical
Sciences, Bilaspur, H.P.


Dr. Muninder Kumar
Additional Professor
Radiation Oncology
AIIMS, Bilaspur (H.P.)
Regd. No. HPNMC-798/09


Dr. Priyanka Thakur
Associate Professor
Radiation Oncology
AIIMS Bilaspur, (HP)-174037
Reg No. HPNMC 1772/2012