

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIIMS-BLS/DENTISTRY/2024/NIQ/07

Dated: 14/12/2024

NOTICE INVITING QUOTATION

Supply of Magnifying loupes for the Department of Dentistry

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for supply Magnifying loupes for the Department of Dentistry as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as "Quotation No. AIIMS-BLS/DENTISTRY/2024/NIQ/07 for the for supply of Equipment and Consumables for the Department of Dentistry due date of opening 29 / 12/2024" containing both the "Technical Bid" and "Price Bid" may be submitted so as to reach on or before 28/12/2024 up to 11:00 A.M in the R.No B 749, Department of Dentistry 4th floor B block AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh. PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee and bidders who may wish to be present on the same day at 11:30 A.M.


For the purpose of technical evaluation, the bidder is required to submit following documents:

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during 2021-22 and 2022-23 for the similar products.
4. Previous year Income Tax Return (AY- 2020-21, 2021-22 and 2022-23).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical Evaluation will be based on total value wise.
8. Bidder must submit sample for each item along with bid as per specification for technical evaluation.

In case the Technical Committee rejects a sample or the bid is "Not Qualified" on technical grounds, the financial bid in respect of all the items will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.



Medical Superintendent
चिकित्सा अधीक्षक
AIIMS, Bilaspur (H.P.)
अभा.आ.स. बिलासपुर (हि.प्र.)

Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other **terms and conditions** will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in Department of Dentistry 4th floor B- Block, AIIMS Bilaspur, Kothipura, Himachal Pradesh PIN- 174001.
13. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact Dr. Dinesh Kumar Verma, Faculty In charge, Department of Dentistry, E-mail office.ms@aiimsbilaspur.edu.in


Faculty In Charge
Department Of Dentistry
AIIMS-Bilaspur
अ.भा.अ.स. (H.P.)
H.P.

Sl. No	Name	Technical specifications (Annexure A)	Qty
1.	Magnifying loupes	<p>The Surgical System should mainly consist of the optical system, adjustable frame and standard accessories</p> <p>OPTICAL SYSTEM:</p> <ol style="list-style-type: none"> 1. Galilean optical system featuring compact design, delivering precise image with good colour fidelity extending to the peripheral zones, excellent depth of field of view ensuring clear visualization of anatomical structures. 2. Magnification: 4.5 X 3. Working Distance (Eye to operating area) of 350 mm or more 4. Field of view diameter at working distance of 350 mm: 55mm or more 5. Quick adjustment of interpupillary distance from 55mm to 80 mm by means of left and right coaxial knobs to suit individual surgeons. 6. High Quality scratch free lens protection device with anti-reflection coating shielding the objective lens against tissue debris for increased protection 7. High quality surface of the optical system resistant to standard disinfectants 8. Telescopic rail enabling quick positioning of the eye piece with a single adjustment. 9. Quick adjustment of eyepiece tilt to desired viewing angles even in extreme treatment options 10. Flip up function for unobstructed vision and eye contact with patients with single adjustment 11. Mount for easy attachment and removal of the optical system 12. Serializable contact guard for reliable swinging of the optical system up and down. 	1

Medical Superintendent
विश्वम्भर अजीसक
AIIMS, Bhaspur (H.P.)
अ.भा.अ.स. (वि.स.स.)

LED LIGHT ILLUMINATION:

1. Should include illumination system attached to the loupes
2. Powerful LED light source with integrated temperature control
3. Should be able to attach easily to the optical system
4. Light intensity up to 50000 Lux
5. Shock proof protection for the light and accessories
6. Light should be resembling day light and should illuminate the whole field with even illumination.
7. Low weight
8. Should include 2 Medical grade Lithium ion rechargeable battery with charging device
9. Should have a run time of above 3 - 4 hrs. at 100 % intensity.
10. Belt clip for the freedom of movement.
11. All the necessary accessories for the proper functioning of the unit and suitable high quality soft case for storage.

HEAD BAND:

1. Sturdy head band with washable textile pads for wearing comfort and easy cleaning.
2. Should be able to adjust to the shape of the operator's head.
3. Should provide even weight distribution on operator's head.
4. Should be able to lock the loupe in the desired viewing angle. **STANDARD ACCESSORIES:** Objective lens protective device (2), Contact Guard (2), cleaning cloth for optical components, Allen Key, accessories for the LED light illumination, High quality soft case for protection of the surgical loupe and accessories.

COMPLIANCE:

1. Should be European CE or US FDA approved
2. Should carry at least 2 years of warranty