



GEM BID REF NO- GEM/2024/B/5466529

ITEM NAME- Laproscopic Instrument Set for Urology Department

Representation from:

1. Accurate Surgical

Representation Received	Remark
<ul style="list-style-type: none">• Working length of all Laparoscopic Instruments except clip applicers should be 300mm to 350mm. Instruments can be with or without leur lock.• Angle variations in Applier -/+ 10 DegreeLength Variation in Applier -/+ 10 Degree• 30 percent same OEM and 20 percent Indian Content/Third Party Reputed Company should be acceptable. There is no bidder can quote 100 percent content of same manufacturer.• Laparoscopic Instruments are advanced surgical instruments and advanced surgical instruments are exempted from MII• Advanced surgical instruments is mentioned in GTE list with serial number 155. We request you please allow the exemption of MII <p>We request you please extend the bid for at least 10 days so that we can participate in the bid.</p>	<p>Accepted</p>

Handwritten signature

Handwritten initials

Dr. Anurag Dhillon
Department of Urology
All India Institute of Medical Sciences (A.P.)
Bilaspur
01762550416



All India Institute of Medical Sciences, Bilaspur
Himachal Pradesh- 174037

We hope our request will be entertained for the maximum bidder's participation in this bid.

Handwritten signature

Handwritten signature
Dr. Bijaya Kumar Sethi
Assistant Professor
Department of General Surgery
AllMS, Bilaspur (H.P.)
Regd. No. ORS1998 0000038KTK

Handwritten signature
Dr. Ajay Kumar Dhimman
Assistant Professor
Department of General Surgery
AllMS, Bilaspur (H.P.)
DMC/65346