

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES**  
**Bilaspur, Himachal Pradesh**

No. AIIMS-BLS/B&PS/2024/NIQ/30

Dated: 15/10/2024

**NOTICE INVITING QUOTATION**

**Supply of 1 set - Microvascular instruments in the Department of Burns and Plastic Surgery**

Quotation as a PASSWORD protected PDF file is invited from registered stockists/ distributors/ manufacturers having GST No. for the Supply of microvascular instruments in the Department of Burns and Plastic Surgery as per the detailed specifications (Annexure-A). Submit quotation as a PASSWORD protected PDF **“Quotation No. AIIMS-BLS/Stores/2023/NIQ/30** for the Supply of microvascular instruments set in the Department of Burns and Plastic Surgery **due date of opening 16/10/2024**” containing both the “Technical Bid” and “Price Bid” (in two separate files) may be submitted so as to reach on or before **05/11/2024** up to 11:00 A.M in Conference hall of Procurement & Stores Department, Basement-1, D-Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee.

For the purpose of technical evaluation, the bidder is required to submit following documents.

**Technical Bid:**

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Turnover for last two years (2021-22 & 2022-23).
4. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
5. Full description, make and quantity of the item must be mentioned in the quotation.
6. Technical Evaluation will be based on total value wise

In case the Technical Committee rejects a sample (if asked for) or the bid is “Not Qualified” on technical grounds, the financial bid in respect of that item will not be considered.

**Financial Bid:** The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail. Should have comprehensive warranty for at least 1 year.

**Financial Bid submission format** for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other terms and conditions will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in Department of burns and plastic surgery, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001.
13. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact Dr. Navneet Sharma, Department of burns and plastic surgery, AIIMS Bilaspur HP 01978-292575, E-mail [dr.navneet.plasticsurgery@aiimsbilaspur.edu.in](mailto:dr.navneet.plasticsurgery@aiimsbilaspur.edu.in)

Head of Department  
Burns and Plastic surgery  
AIIMS-Bilaspur  
Medical Superintendent  
H.P.  
चिकित्सा अधीक्षक  
AIIMS, Bilaspur (H.P.)  
अ.भा.आ.स. बिलासपुर (हि.प्र.)

### Microvascular Instruments' Specifications

- Instruments should be made from High-Quality Surgical Grade Stainless Steel or Titanium (or as otherwise wherever specified individually) and should be rust-proof, with excellent non-reflective, superior grip finish, and special ergonomic design for ease of handling.
- Instruments should have a light and compact design.
- Instruments should be extremely durable.
- Instruments should have a unique gripping surface
- Instruments should have at least 1-year manufacturer's warranty for any manufacturing defects.
- Instruments should have the following specifications:

Microvascular Customized instrument Set									
Specification, Sizes, and Qty									
S.No	Instrument Name	Handle Type	Working Length Type	Width of Handle (mm)	Tip Size (mm)	Opening (mm)	Total Length (mm)	Material	QTY
1	Microvascular Scissor	Round Handle	Straight	8	0.4	4	150	Ti/ss	02
2			Curved	8	0.4	4	150		01
3	Needle Holder	Round Handle	Straight	8	0.6	4	150	Ti/ss	01
4			Curved	8	0.6	4	150		01
5	Vessel Dilators	Flat Handle	Straight Tip		0.4	6	110	Ti/ss	01
6					0.4	6	130		01
7	Jeweller Forceps	Flat Handle					110		01
8							120		02
9							130		02
10		Angulated Flat Handle				0.3	7	130	02
11	Jeweller Forceps	Toothed Flat Handle				9	130		02
12	Instrument Box with Silicon Mat								02

*Medical Superintendent*

**Dr. Chitresh Kumar**  
Associate Professor  
Dept. of Surgical Oncology  
AIIMS, Jhansi (H.P.)  
Phone No. 0511-26278

S.No	Clamp	Size	Qty	Blood Vessel Dia. Range (mm)	Calibrated Pressure (gm)	Jaw Width (mm)	Clamp Length (mm)	Material
13	2 Jaw Microvascular Clamps, Venous, Tandem	3VV	1 Pair	1.0 - 2.25	42~45	1.7	16	Stainless Steel
14	2 Jaw Microvascular Clamps, Artery, Tandem	2A	1 Pair	1.0 - 2.25	42-45	1.7	16	Stainless Steel

*[Faint handwritten signature and stamp]*

**Dr. Chitresh Kumar**  
 Associate Professor  
 Dept. of Microbiology  
 AllMS, Siliguri (H.P.)  
 Regd. No. UPRC 51678

**Mr. Anupam Kumar**  
 Superintendent  
 AllMS, Siliguri (H.P.)  
 Regd. No. UPRC 51678