

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh**

No. AIIMS-BLS/Surg.Onco/2025/NIQ/02

Dated: 07/03/25

NOTICE INVITING QUOTATION

Supply of Point of care Testing Equipment Coagulation Analyzer under miscellaneous items for the department of Surgical Oncology

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for the Supply of **Point of care Testing Equipment Coagulation Analyzer under miscellaneous items** to the Department of Surgical Oncology as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as "**Quotation No. AIIMS-BLS/Surg.Onco/2025/NIQ/02** for the Supply of ATP Monitoring Luminometer under miscellaneous items Department of Surgical Oncology **due date of opening 22/03/25**" containing both the "Technical Bid" and "Price Bid" (in two separate envelopes) may be submitted so as to reach on or before 21/04/25 up to 11:00 A.M in the room no.B-716 Department of Surgical Oncology 4th floor B- Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee .

For the purpose of technical evaluation, the bidder is required to submit the following documents.

Technical Bid:

8. Profile-Name & Full Address of the firm & year of establishment.
9. Copy of permanent GST registration certificate.
10. Copies of supply orders secured during 2023-24 for the similar products.
11. Turnover for last two years (2021-22 , 2022-23 & 2023-24).
12. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
13. Full description, make and quantity of the item must be mentioned in the quotation.
14. Technical Evaluation will be based on total value wise

In case the Technical Committee rejects a sample (if asked for) or the bid is "Not Qualified" on technical grounds, the financial bid in respect of that item will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.


Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

(Handwritten signatures and initials)

Other terms and conditions will be as follows: -

14. Supply will be F.O.R. at AIIMS-Bilaspur
15. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
16. Supply must conform to samples wherever asked for.
17. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
18. Samples will be preserved till final settlement.
19. Bidder is required to quote rate of 100% items.
20. Part supply will not be accepted.
21. No revision in rate (on higher side) will be accepted at any stage.
22. The firm shall not assign or sublet the work/job or any part of it to any other firm.
23. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected and accepted. No advance payment will be made at any stage.
24. Taxes at other government levies will be paid extra as applicable.
25. Venue for sample verification (if asked for) and quotation opening shall be in room no.B-716 Department of Surgical Oncology 4th floor B- Block, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. In case manufacturer participates, dealers will not be allowed or disqualified.
26. For any query, please contact Dr. Chitresh Kumar Sharma, Faculty In charge, department of SurgicalOncology 01978-292575, E-mail :-office.surgon@aiimsbilaspur.edu.in


Faculty In Charge
Dr. CHITRESH KUMAR
Assistant Professor
Department of Surgical Oncology
AIIMS-Bilaspur
H.P. ASPUR (H.P)
Regd. No. UPMC-51678

ANNEXURE- A

Technical Specifications for Coagulation analyzer for PT/INR testing:-

Sample type:-Finger prick of blood / venous blood,Sample volume:- 8 μ L
Haematocrit range:- 25 — 55% inclusive,Reporting range :-0.8 — 8.0 INR
Results displayed 'International Normalized Ratio' (INR) and PT
in seconds

Features :-Automatic power off Configurable: disabled, 5 mins, 15 mins, 30
mins,Dimensions :-147 \times 84 \times 32 mm (5.79 \times 3.31 \times 1.26 in)

Weight 210 g :-(7.41 oz),Memory recall 1000 patient results

Design lifetime 3 years or 8,760 tests

Meter operating conditions,Temperature 15°C — 32°C (59°F — 89°F)

Relative humidity < 80% (without condensation)

Altitude Electrical safety tested up to 2000 m

Transport and storage conditions (in packaging)

Temperature -20°C — 40°C (-4°F — 104°F)

Relative humidity < 75% (without condensation)

Altitude Electrical safety tested up to 2000 m

Test strips storage conditions

Temperature 2°C — 30°C (35.6°F — 86°F)

Relative humidity :-< 75% (without condensation)

LQC kit storage conditions :-Temperature 2°C — 8°C (35.6°F — 46.4°F)

Power supply

Model GTM46161-165.0-USB,Input :-100 — 240 V~, 50 — 60 Hz, 0.45 A

Mains supply voltage fluctuation +/-10%,Rating 5V 3.2A

Operating altitude up to 5000 m,Operating humidity 0 — 93%

Operating air pressure 54 — 1060 hPa,Operating temperature 0°C — 40°C (32°F —
104°F),Connectivity:-USB, Micro USB, WiFi 802.11b/g/n

PC requirements,Hardware,WiFi adapter (for WiFi connection)

Device must be connected to the same wireless network as the analyzer.

Spare USB port (for USB connection),Software Web browser

Internet Explorer is not supported

General:-Input 5V 0.9A,Rating 3.7V 0.5A MAX

Capacity Rechargeable 1.85Ah (LI-ION) with power monitoring

Water ingress rating IPX0

Pollution degree rating Pollution degree II

Overvoltage category Category I

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]