

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIIMS-BLS/OBG/2025/NIQ/IUI/02

Dated: 15-01-2026

NOTICE INVITING QUOTATION
Supply of IUI Items

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for the Supply of IUI items in the Department of Obstetrics and Gynaecology as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as “**Quotation No. AIIMS-BLS/OBG/2025/NIQ/IUI/02** for the Supply of IUI items in the Department of Obstetrics and Gynaecology due **date of opening 31-01-2026**” containing both the “Technical Bid” and “Price Bid” (in two separate envelopes) may be submitted so as to reach on or before **30-01-2026** up to 11:00 A.M in Room No - 535, OPD Block B, 2nd floor, Department of Obstetrics and Gynaecology, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee.

For the purpose of technical evaluation, the bidder is required to submit following documents.

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of PAN, GST registration certificate.
3. Copies of previous supply orders secured during 2024-25 for the similar products.
4. Turnover for last two years (2023-24 & 2024-25).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical Evaluation will be based on total value wise

In case the Technical Committee rejects a sample (if asked for) or the bid is “Not Qualified” on technical grounds, the financial bid in respect of that item will not be considered.


Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.

Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other terms and conditions will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in Room No - 535, OPD Block B, 2nd floor, Department of Obstetrics and Gynaecology, AIIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh PIN- 174001.
13. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact Dr.Sushruti Kaushal, Associate Professor, AIIMS Bilaspur, Contact no 9805657709 email id dr.sushruti.obg.aiimsbilaspur.edu.in


-sd-
Dr. Poojan Dogra Marwaha
Professor & HOD, Department of Obstetrics & Gynaecology, AIIMS Bilaspur
AIIMS Bilaspur (H.P) - 174001
अखिल भारतीय आयुर्विज्ञान संस्थान, बिलासपुर, हिमाचल प्रदेश
HPMC 4124/19 / एचपीएमसी 4124/19

Annexure-A

Specifications of Items

Sr. No	Items Name	Quantity	Specifications
01	Centrifuge with swing out rotor 4 x 15 (refrigerated)	01	<p>With refrigerated Centrifuge Medical Institutional laboratories available with a wide choice of rotor heads and adaptors, Should be the basic unit with speed regulator.</p> <p>Stepless speed regulator.</p> <p>Safety Lid interlock to prevent cover opening during centrifugation.</p> <p>Digital speed meter and 0-60 minutes digital countdown timer. It should have a timer and a rotor (upto 3000rpm) whose speed can be controlled .</p>
02	Vertical Laminar Flow	01	<p>Stainless Steel Table top</p> <p>Leveling Jacks with</p> <p>HEPA Filter (0.3 Microns)</p> <p>Power Sockets : 3 Nos. of 5 Amps Sockets</p> <p>Power Supply : 220 V Single Phase , 50 Hz. 1.2 KVA</p> <p>Temp. accuracy : $\pm 0.3^{\circ}\text{C}$ (Room Temp needs to maintain 9°C less than set pt.)</p> <p>Temp. Uniformity : $95\% \pm 5\%$</p> <p>HEPA- Filter</p> <p>Media : Ultra clean glass Fiber paper - Imported</p> <p>Casing : Aluminium Box Type</p> <p>Separators : Minipleat no separators</p> <p>Retention : 0.3 Microns and above</p> <p>Efficiency : 99.997 %</p> <p>Pre- Filter</p> <p>Media : Synthetic, Non-woven polyester fibers</p> <p>Casing : Aluminium Flange Type</p> <p>Retention : 10 Microns</p> <p>Efficiency : 99.97 %</p> <p>INTEGRATED TABLE HEATING SYSTEM</p> <p>1. Integrated Table Heating area 45 Cms x 50 Cms. With</p>

			<ul style="list-style-type: none">2. Microprocessor Based Temperature controller3. Heating Coil for Uniformity & Stability STAGE4. Microscope Fittings with Heating Stage Warmer with Microprocessor Based Temperature controller -
--	--	--	---

Handwritten blue ink scribbles and checkmarks below the table.

s