



अखिल भारतीय आयुर्विज्ञान संस्थान , बिलासपुर  
हिमाचल प्रदेश -174001

**All India Institute of Medical Sciences, Bilaspur  
Himachal Pradesh-174001**

**Website: [www.aiimsbilaspur.edu.in](http://www.aiimsbilaspur.edu.in)  
email: [storeofficer@aiimsbilaspur.edu.in](mailto:storeofficer@aiimsbilaspur.edu.in)**



**Corrigendum 2**

**Tender Reference No:** AIIMS/BLS/G/2024-24/E-Tender/13

**Date:** 27/03/2025

**Tender ID:**2025\_AMSBL\_852976\_1

**Supply of Consumables for PET/CT Nuclear Medicine Department on Rate Contract Basis.**

Reference to the above- mentioned Tender ID “**Supply of Consumables for PET/CT Nuclear Medicine Department on Rate Contract Basis**” uploaded on CPPP on dated 13/03/2025 with end date 09/04/2025.

All prospective Bidders are hereby informed that the following terms and conditions have been changed in above said tender:

| S.No | Raised clarifications   | Modified Terms and conditions  |
|------|---|--|
| 1    | AERB Procurement NOC is not mentioned in the Tender. What if the Purchaser does not have NOC? | NOC will be provided by the Department of Nuclear Medicine to the successful bidder.   |
| 2    | Quality Documents like ISO, etc. - Only Quality report we can share                           | It is acceptable. QA/QC reports should include Radiochemical and Radionuclide purity, apyrogenicity, sterility, chemical purity and certificate of safety for IV administration in human patients. |
| 3    | Clause 6.1— How the delivery order will be made?  | PO will be issued on weekly basis with staggered delivery.   |
| 4    | Clause 7— Warranty is not applicable on radio pharmaceuticals.                                | Acceptable. QA/QC reports should be furnished instead.   |
| 5    | Clause 25: Supply Time — 9AM?? Our product will be delivered as per calibration time.         | Typographical error. Time will be as per Schedule 1 of technical specifications.   |
| 6    | Clause 28 - There should be enough time for supply as per AERB limit of Purchaser.            | Agreed.  |
| 7    | Clause 32— Shelf life not applicable  | Agreed. Compliance with calibration time will be used instead.   |
| 8    | Sample providing is not possible  | Proof of supply to a running PET/CT center should be provided instead.   |
| 9    | Product Specification: we can share quality control reports.                                  | Agreed. Refer to point 2 above.  |

The other terms and conditions of the tender shall remain same.

*Ad*  
27/3/25

F.I. Procurement