

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No-AIIMS-BLS/G/2025/General item/NIQ/383

Dated: 22/02/2025

NOTICE INVITING QUOTATION
Supply of Laryngoscope cells

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for the Supply of Laryngoscope cells as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as “**Quotation No. AIIMS-BLS/G/2025/General item/NIQ/383 for the Supply of Stamps due date of opening 08/04/2025**” containing both the “Technical Bid” and “Price Bid” (in two separate envelops) may be submitted so as to reach on or before 08/04/2025 up to 11:00 A.M in Procurement Section, E block, Ground floor, AIIMS Bilaspur, Kothipura, Himachal Pradesh PIN- 174001. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee.

For the purpose of technical evaluation, the bidder is required to submit following documents.

Technical Bid:

1. Pan card of the owner of the business.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders or bills secured during 2022-2023 & 2023-24 for the similar products.
4. Turnover for last two years (2022-23 & 2023-24).
5. Self-declaration that the firm is not debarred by any government organisation or PSUs.

The bidder should sign and stamp each page of quotation document as a token of acceptance of the terms & conditions contained therein and submit the same along with the bid.

In case the Technical Committee rejects a sample (if asked for) or the bid is “Not Qualified” on technical grounds, the financial bid in respect of that item will not be considered.

Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for 3 months. In case discrepancy between unit price & total price, the unit price shall prevail.


Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item	Specifications	Quantity	Unit Price	GST %	Unit price with GST	Total price
1.	Laryngoscope	Attached as Annexure -A	400 units				

5

Other terms and conditions will be as follows: -

1. Supply will be F.O.R. at AIIMS Bilaspur H.P.
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. No revision in rate (on higher side) will be accepted at any stage.
8. The firm shall not assign or sublet the work/job or any part of it to any other firm.
9. Billing will be in the name of Executive Director, AIIMS Bilaspur H.P.
10. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes & other government levies will be paid extra as applicable.
12. In case manufacturer participates, dealers will not be allowed or disqualified.
13. For any query, please contact through E-mail storeofficer@aiimsbilaspur.edu.in



Faculty In-Charge
(Procurement)
AIIMS Bilaspur H.P.

SPECIFICATION FOR TYPE C BATTERY (LARYNGOSCOPE BATTERY)

Sr No.	Specification	Details
1.	Battery Type	Zinc-Carbon (Leclanché chemistry)
2.	Model	R14 (UM-2U), Type C
3.	Voltage	1.5 Volts
4.	Size	Type C (Medium cylindrical battery)
5.	Dimensions: Diameter: Height:	25.0–26.2 mm 49.5–50.5 mm
6.	Weight	Approx. 60–70 grams
7.	Nominal Capacity	3,000 to 8,000 mAh (varies with load and discharge rate)
8.	Operating Temperature	-10°C to 45°C (optimal performance: 0°C to 25°C)
9.	Shelf Life	2 to 3 years (under standard storage conditions)
10.	Terminal Type	Flat positive terminal with a protruding nub and flat negative base
11.	Discharge Characteristics	Continuous: Low to medium current; Intermittent: Handles peak loads
12.	Construction	Outer Jacket: Steel can or plastic cover
13.	Electrolyte	Ammonium chloride and zinc chloride
14.	Core	Carbon rod with manganese dioxide (positive electrode)
15.	Environmental Considerations	Non-rechargeable; dispose of properly. No mercury or cadmium.
16.	Compliance Standards	IEC R14, ANSI C

N/O
25/1/24.

Dr. Tarun Sharma
Associate Professor
Internal Medicine
AIIMS Bilaspur (H.P.)
HPMC-1342/2011