

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Bilaspur, Himachal Pradesh

No. AIMS-BLS/DENTISTRY/2024/NIQ/08

Dated: 17/01/2025

NOTICE INVITING QUOTATION

Supply of Miscellaneous Items for the Department of Dentistry

Sealed quotations are invited from registered stockists/ distributors/ manufacturers having GST No. for supply Miscellaneous Items for the Department of Dentistry as per the detailed specifications (Annexure-A). Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as "**Quotation No. AIMS-BLS/DENTISTRY/2024/NIQ/08**" for the for supply of Equipment and Consumables for the Department of Dentistry **due date of opening 02/02/2025**" containing both the "Technical Bid" and "Price Bid" may be submitted so as to reach on or before 01/02/2025 up to 11:00 A.M in the R.No B 749, Department of Dentistry 4th floor B block AIMS Bilaspur, Kothipura, Bilaspur, Himachal Pradesh. PIN- 174001. Bidders have to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted local purchase committee and bidders who may wish to be present on the same day at 11:30 A.M.

For the purpose of technical evaluation, the bidder is required to submit following documents:

Technical Bid:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during 2021-22 and 2022-23 for the similar products.
4. Previous year Income Tax Return (AY- 2020-21, 2021-22 and 2022-23).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical Evaluation will be based on total item wise. Vendor should quote for all the items.
8. Bidder must submit sample for each item along with bid as per specification for technical evaluation.

In case the Technical Committee rejects a sample or the bid is "Not Qualified" on technical grounds, the financial bid in respect of all the items will not be considered.


Financial Bid: The rates and total cost must be quoted in both words and figures (over writing not allowed). Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.

Financial Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty Required	Make	Model	HSN Code	Unit Price	GST @	Unit price with tax	Total Amount

Other **terms and conditions** will be as follows: -

1. Supply will be F.O.R. at AIIMS-Bilaspur
2. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
3. Supply must conform to samples wherever asked for.
4. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within two weeks' period. The Institute will not bear any expenses on this account.
5. Samples will be preserved till final settlement.
6. Bidder is required to quote rate of 100% items.
7. Part supply will not be accepted.
8. No revision in rate (on higher side) will be accepted at any stage.
9. The firm shall not assign or sublet the work/job or any part of it to any other firm.
10. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected & accepted. No advance payment will be made at any stage.
11. Taxes at other government levies will be paid extra as applicable.
12. Venue for sample verification (if asked for) and quotation opening shall be in Department of Dentistry 4th floor B- Block, AIIMS Bilaspur, Kothipura, Himachal Pradesh PIN- 174001.
13. In case manufacturer participates, dealers will not be allowed or disqualified.
14. For any query, please contact Dr. Dinesh Kumar Verma, Faculty In charge, Department of Dentistry, E-mail office.ms@aiimsbilaspur.edu.in


Faculty In Charge
Department Of Dentistry
AIIMS-Bilaspur
H.P.

Medical Superintendent

चिकित्सा अधीक्षक
AIIMS, Bilaspur (H.P.)
अ.भा.आ.स. बिलासपुर (हि.प्र.)

Annexure A

Department of Dentistry

Sl. No	Name	Technical specifications	Quantity
1.	Heavy duty dental vibrator	<ol style="list-style-type: none"> 1. Should have capacity of 3-4 casting cylinders. 2. Should have removable rubber plate for cleaning. 3. Should be fitted with speed controller regulator. 4. Should have heavy duty solid base made up of rust proof material. 5. Sound proof continuous vibration. 	01
2.	Electric pulp tester	<ol style="list-style-type: none"> 1. Automatic one-button operation 2. Three rates of stimulus increase - slow, medium and fast 3. Memory of the previous rate of speed programmed 4. Micro-processor based self-test assuring proper function of the unit every time 5. Display should remain frozen for about 3 minutes after releasing the button. 6. With 4 autoclavable probes: standard long, standard short, labial, lingual 7. 9-V alkaline battery. 8. One year power unit warranty. 9. Size: Gentle Pulse: 178 mm (+5 mm) x 22 mm (+2mm) x 32 mm (#2 mm) 10. Weight: 120±10 gm, Gentle Pulse 11. Length: Ground Wire Assembly: 30" (76cm) 12. Battery: Standard 9-V 	01
3.	Model trimmer	<ol style="list-style-type: none"> 1. Non rusting body 2. Both wet and dry trimming should be possible. 3. Housing should be sound dampening and resistant. 4. Speed should be between 2500-3000 Rpm. 5. Adjustable work table 	01

		<ol style="list-style-type: none"> 6. An orthodontic angulator: platform for holding models for accurate trimming with different angulation. 7. Valve for regulating the water flow, supplied with water tubing. 8. Should be supplied with all accessories to make the unit functional. 9. Should be European CE compliant. 	
4.	Water Distiller	<ol style="list-style-type: none"> 1. Should be European CE with 4 digit notified body no. or US FDA approved 2. The boil chamber and the condensing coils should be made of at least 304 stainless steels. 3. Should have a working efficiency of at least 1 litre of distilled water in 1 hour. 4. It should have high quality plastic collection jar. 5. The removable lid and handle of the jar should be of food grade polypropylene. 6. It should have an automatic shut off at the end of a cycle. 	01