



GEM BID REF NO-GEM/2024/8/5542027

ITEM NAME- Open Surgical Instrument for Renal Transplant for Department of Urology

Representation received	Remark
<p>1) instruments and Containers should be from same manufacturer.</p> <p>The permanent filter is validated for 5,000 cleaning and sterilization cycles (for steam sterilization through fractionated vacuum process validated according to DIN EN ISO 17665 part 1).</p> <p>2) Manufacturer details must be mentioned on Container lid and base.</p> <p>3) Locking flaps of the lids can be opened outwards Locking flaps can be arrested in any position (up to 180) and folded back to the lid so that flaps do not get caught during the washing process</p> <p>4) The Container must European CE MDR Certified/CDSCO/BIS certified and certificate should be attached with the bid.</p> <p>5) Physical demonstration of containers is must during technical evaluation of bids.</p> <p>6, Renal Transplant Instruments are advanced surgical instruments and advanced surgical instruments are exempted from MII. Advanced surgical instruments is mentioned in GTE list with serial number 155. We request you please allow the exemption MII</p>	<p>Accepted</p>

L/2024

Dr. Saumya Chopra
MBBS, MS General Surgery
Assistant Professor
Department of General Surgery
AIIMS, Bilaspur
DMC/R/07953

Dr. Bijay Kumar Sethi
Associate Professor
Deptt. of Plastic Surgery
AIIMS, Bilaspur
Regd. No. QRS1300000006X TK